

=> d his

(FILE 'HOME' ENTERED AT 13:44:28 ON 23 APR 2003)

FILE 'CA' ENTERED AT 13:44:34 ON 23 APR 2003

L1 52 S (PIPET? OR DISPENS? OR DILUT?) (5A) (NANOLITER OR NL)  
L2 36 S L1 NOT PY>2000  
L3 5 S L1 NOT L2 AND PATENT/DT  
L4 41 S L2-3

=> d bib,ab 1-41

L4 ANSWER 9 OF 41 CA COPYRIGHT 2003 ACS

AN 133:249314 CA

TI Microreactor systems and methods for performing reactions in an unsealed environment

IN Becker, Thomas; Koster, Hubert; Cantor, Charles R.

PA Sequenom, Inc., USA

SO PCT Int. Appl., 95 pp.

PI WO 2000056446 A1 20000928 WO 2000-US6288 20000310  
US 6225061 B1 20010501 US 1999-266409 19990310

PRAI US 1999-266409 A1 19990310

AB An open microreactor system is described for performing a sub-microliter reaction. The open system can contain a solid support having a target site for performing the reaction; a liq. dispensing system such as a nanoliter dispensing pipet for dispensing a sub-microliter amt. of a liq. to the target site; a temp. control device for regulating the temp. of the support; and means for controlling the amt. of liq. dispensed, which corresponds to the amt. of liq. that evaps. from the target site. The support can be a (functionalized) bead, pin, comb, wafer, well or microchip. The reaction can include nucleic acid amplification, combinatorial library synthesis, biopolymer sequencing or primer oligo base extension (PROBE).

L4 ANSWER 10 OF 41 CA COPYRIGHT 2003 ACS

AN 132:331472 CA

TI Implementation of nanoliter dispensing in the laboratory

AU Bulow, Sven

CS Eppendorf-Netheler-Hinz GmbH, Hamburg, D-22339, Germany

SO GIT Labor-Fachzeitschrift (2000), 44(4), 396,398-399

LA German

AB The nL dispenser Nanozyme is described for a reliable and reproducible dosage of vols.  $\geq 10$  nL giving its principle and main application fields.

L4 ANSWER 14 OF 41 CA COPYRIGHT 2003 ACS

AN 131:103851 CA

TI Apparatus for dispensing a predetermined volume of a liquid

IN Backes, Monica; Corless, Anthony Robert; Shaw, John Edward Andrew; Sibbald, Alastair

PA Central Research Laboratories Limited, UK

SO PCT Int. Appl., 20 pp.

PI WO 9936176 A1 19990722 WO 1999-GB163 19990118

PRAI GB 1998-933 A 19980117

AB A dispensing app. is described for dispensing a predetd. vol. (1 nl to 2  $\mu$ l) of a liq., e.g., liq. biochem. reagents, and comprises a liq. reservoir, a channel with an outlet for conveying the liq. from the reservoir to the outlet, and means for generating a pulse of gas. The flow of gas causes a predetd. vol. of liq. to be ejected from the outlet. The outlet comprises a pair of openings in the channel which face one another,

the liq. being retained between the openings by surface tension in the absence of a flow of gas. The gas flow is directed towards one of the openings. The app. avoids contamination of the liq., and can be produced as a low cost disposable unit. The app. is less sensitive to liq. viscosity than existing devices.

L4 ANSWER 16 OF 41 CA COPYRIGHT 2003 ACS  
AN 130:47042 CA  
TI MultiPROBE nL complements drug discovery assay miniaturization  
AU Driscoll, Jennifer; Delmendo, Ron; Papen, Roeland; Sawutz, David  
CS Small Molecule Chemistry, Amgen, Inc., Thousand Oaks, CA, 91320, USA  
SO Journal of Biomolecular Screening (1998), 3(3), 237-239  
AB The Packard MultiPROBE nL is designed to enable the MultiPROBE Automated Liq. Handling System to aspirate and dispense nanoliter vols. Several features add confidence to small vol. transfers. A preview of nanoliter dispensing can be seen on a video camera monitor. In addn. to the std. wash station, syringe and ultrasonic flushes can be run at the start of a program to prevent dirt or air obstructions. The MultiPROBE nL can transfer ionic, nonionic, and solns. contg. org. solvents such as DMSO directly from master to assay plates and into high-d. plate arrays. Addnl., the MultiPROBE nL increases the efficiency of generating dose response curves for secondary screening by eliminating a diln. step. IC50 values obtained after compd. prepns. with the instrument are consistent with those values previously detd. using an MultiPROBE 208.

L4 ANSWER 20 OF 41 CA COPYRIGHT 2003 ACS  
AN 128:272123 CA  
TI A pneumatically actuated micropipetting device  
AU Szita, Nicolas; Buser, Rudolf  
CS Institute of Mechanics, ETH Zurich, Zurich, CH-8092, Switz.  
SO Proceedings of SPIE-The International Society for Optical Engineering (1998), 3258(Micro- and Nanofabricated Structures and Devices for Biomedical Environmental Applications), 156-163  
AB A valveless micropipetting device is realized with an integrated sensor which can aspirate and dispense liq. vols. without any valves, hence without any reflow or dead vol. With an external pneumatic actuation, aspirating and dispensing are demonstrated from 190 nL -6  $\mu$ L of water. Measurements showed a std. deviation of down to 1%. An integrated capacitive sensor allows monitoring of the pressure throughout the pipetting process and detect malfunctions, e.g. clotting of the pipetting tip. The aspiration mechanism is used in combination with a micromachined reaction chamber and a miniaturized optical anal. system.

L4 ANSWER 28 OF 41 CA COPYRIGHT 2003 ACS  
AN 109:66130 CA  
TI Pneumatic microsyringe for use as an injector in open tubular liquid chromatography and as a dispenser in microanalysis  
AU Kennedy, Robert T.; Jorgenson, James W.  
CS Dep. Chem., Univ. North Carolina, Chapel Hill, NC, 27599-3290, USA  
SO Analytical Chemistry (1988), 60(15), 1521-4  
AB A pneumatic microsyringe is described and characterized for use as a microinjector for open tubular liq. chromatog. and for use as a microdispenser. As a microinjector, the syringe had a relative std. deviation of 2.7% in vol. delivered for 10-nL injections on a 15  $\mu$ m inner diam. column and showed the same contribution to peak broadening as other methods of injection. The vol. injected is easily changed by simply changing the length of time the injection is made. This device is useful for injecting samples of limited vol. As a microdispenser, the device had

a relative std. deviation of 3.38% in vol. delivered for dispensing 0.248 nL. In this fashion the syringe can be used to add internal stds. or reagents to small samples. The syringe can be calibrated for microdispensing by measuring the size of a droplet formed from injecting an aq. soln. into mineral oil.

L4 ANSWER 29 OF 41 CA COPYRIGHT 2003 ACS  
AN 108:189019 CA  
TI Piston stroke pipet for nanoliter delivery  
IN Koeppen, Bernd; Neymeyer, Hans Georg; Quandt, Ingeborg; Moeschwitzer, Gerhard  
PA Ministerium des Innern, Ger. Dem. Rep.  
SO Ger. (East), 2 pp.  
PI DD 251086 A1 19871104 DD 1986-292579 19860717  
PRAI DD 1986-292579 19860717  
AB The title pipet for delivery of fluids in the nL range has a 3-stroke operation where the 1st stroke fixes and delivers the fluid vol., the over stroke expels the residual liq., and the after stroke forms an air cushion in the tip for preventing contamination of the fluid yet present in the tip.

L4 ANSWER 30 OF 41 CA COPYRIGHT 2003 ACS  
AN 100:82289 CA  
TI Method for the preparation of biological fluids, at volumes of less than a nanoliter, for their quantitative electron probe analysis  
AU Roinel, N.  
CS Dep. Biol., CEN Saclay, Gif-sur-Yvette, 91191, Fr.  
SO Microanal. X Biol. (1983), 133-9. Editor(s): Quintana, Carmen; Halpern, Sylvain. Publisher: Soc. Fr. Microsc. Electron., Paris, Fr.  
LA French  
AB The prepn. of biol. fluids for element detn. by x-ray microanal. is discussed with respect to pipets for delivering vols. <1 nL, the prepn. of std. solns., prepn. and deposit of samples, and conditions for anal., and examples are given of the prepn. of a std. soln. for the detn. of various elements in mammalian kidney and of std. curves for the detn. of elements in mammalian plasma.

L4 ANSWER 32 OF 41 CA COPYRIGHT 2003 ACS  
AN 99:154575 CA  
TI New techniques and tools for clinical chemistry  
AU Hieftje, Gary M.  
CS Dep. Chem., Indiana Univ., Bloomington, IN, 47405, USA  
SO Clinical Chemistry (Washington, DC, United States) (1983), 29(9), 1659-64  
AB Several new tools of potential use in clin. chem. are described and evaluated. The first, intended to minimize required sample vols., is a device with which a total sample vol. of 1 μL can be dispensed in the form of 1000 identical aliquots. Any no. of such nanoliter aliquots can be taken if larger samples are needed. The 2nd new tool is one for detecting anions or cations sepd. by ion chromatog. Unlike conventional conductometric detectors used in ion chromatog., the new system offers potential sensitivities in the submicrogram/L range and useful operating ranges up to 100 mg/L. The 3rd tool is a scheme for background correction in at. absorption spectrometry; the new technique requires no special auxiliary sources or double-beam optics. Finally, fluorescence time-decay curves and fluorescence lifetimes are shown to be able to overcome the effects of diffusional quenching and scattering resulting from turbidity of solns. in clin. fluorometry.

L4 ANSWER 40 OF 41 CA COPYRIGHT 2003 ACS  
AN 67:50890 CA  
TI Handling microliter and submicroliter quantities of solutions  
AU Sanz, Manuel C.  
SO Memoirs of the Society for Endocrinology (1967), No. 16, 27-35  
AB The actual state of instrumentation for the quant. handling of  
submicroliter quantities is discussed. A new nanoliter pipette is  
described. Procedures for pipette calibration and for correct manipulation  
are given.

L4 ANSWER 41 OF 41 CA COPYRIGHT 2003 ACS  
AN 62:62591 CA  
OREF 62:11123b-d  
TI Constant volume, self-filling nanoliter pipet --construction and  
calibration  
AU Prager, Denis J.; Bowman, Robert L.; Vurek, Gerald G.  
CS Natl. Heart Inst., Bethesda, MD  
SO Science (1965), 147(3658), 606-8  
AB Pipets ranging from <1 to 200 nanoliters can be constructed by a fairly  
simple mech. procedure consisting of sealing, by fusion, a short length of  
small-bore quartz tubing into a long piece of soft glass support-tubing of  
larger bore. When the tip of the pipet is introduced into fluid, the  
quartz tubing fills automatically and completely by capillary action.  
Application of pressure by a syringe delivers a known vol. of fluid with a  
repeatability of 1%. The pipets are calibrated either by radioactive or  
fluorescence techniques, and are esp. useful for transferring biol. fluids.

=> log y  
STN INTERNATIONAL LOGOFF AT 13:47:20 ON 23 APR 2003